

Results Table

If the test was conducted successfully, your result will be one of two approximate value ranges shown in the table below.

1 BDL Result is below detection level if...



The red line on the **left** (next to #1) is **DARKER** than the red line on the right (next to #2).

OR



Only the red line on the **left** (next to #1) is visible.

Indication: Lead concentration is below detection level of 3.8 parts-per-billion.

2 ADL Result is at or above detection level if...



The red line on the **right** (next to #2) is **DARKER** than the red line on the left (next to #1).

OR



Both red lines are **EQUALLY DARK**.

Indication: Lead concentration is at or above detection level of 3.8 parts-per-billion.

If no lines appear on the test strip, your test result is inconclusive. For questions about inconclusive results or having your water laboratory-tested, please see the included full-length instructions sheet.