

## CLIENT INFORMATION

Client: \*\*\*\*\*

Requested On: Dec 4, 2025

Phone: \*\*\*\*\*

Email: \*\*\*\*\*

Kitting, Logistics, and Support provided by: SimpleLab, Inc.

Questions? For fastest assistance:

[support@mytapscore.com](mailto:support@mytapscore.com)

Do not contact facility technicians directly.

## TESTING PERFORMED

**Testing Requested:** Advanced Spring Water Test

**Matrix:** Drinking Water

**Testing / Report ID:** EQLEGA

**Testing Facility:** Microbac Laboratories

**Facility Location:** 428 Route 315 Hwy  
Pittston, Pennsylvania 18640

## SAMPLE INFORMATION

**Collection Date:** Dec 9, 2025

**Collected By:** \*\*\*\*\*

**Received Date:** Dec 11, 2025

**Reported On:** Dec 18, 2025

**Sample Location:** Spring

**Sample Address:** \*\*\*\*\*

## TESTING NOTES FROM TAP SCORE

There were no problems with analytical events associated with this report unless noted. Quality control data is within laboratory defined or method specified acceptance limits except where noted. If you have any questions regarding these test results, please contact [support@mytapscore.com](mailto:support@mytapscore.com)

## SUMMARY ANALYSIS

ANALYTE	UNIT	RESULT	METHOD	EVALUATION
pH	pH	<b>6.7</b>	SM 4500-H+B	OK
Total Dissolved Solids	mg/L	<b>23.5</b>	SM 2510 B	No benchmark available
Turbidity	NTU	<b>0.6</b>	EPA 180.1	No benchmark available
Hardness	mg/L	<b>13.1</b>	SM 2340 B	No benchmark available
Hardness (Total)	mg/L	<b>13.15</b>		
Grains per gallon	Grains	<b>0.77</b>		
Alkalinity (as CaCO <sub>3</sub> )	mg/L	Below Reporting Limit	SM 2320 B	No evaluation possible
Sodium Adsorption Ratio		<b>0.18</b>		

## TEST RESULTS

ANALYTE	UNIT	RESULT	MDL	RL	METHOD	EVALUATION
Aluminum	mg/L	Below Reporting Limit	0.101	0.2	EPA 200.7	No evaluation possible
Antimony	mg/L	Below Reporting Limit	3.0E-5	0.0004	EPA 200.8	No evaluation possible
Arsenic	mg/L	Below Reporting Limit	0.00015	0.001	EPA 200.8	No evaluation possible
Barium	mg/L	<b>0.00394</b>	4.0E-5	0.0004	EPA 200.8	< HGL
Beryllium	mg/L	Below Reporting Limit	4.0E-5	0.0004	EPA 200.8	No evaluation possible
Bicarbonate	mg/L	N/A			Bicarbonate	No benchmark available
Boron	mg/L	Below Reporting Limit	0.0131	0.03	EPA 200.7	No evaluation possible

*Use of this report can only be done in full, with no alterations or additions. SimpleLab maintains the right to enforce this and by accepting this report you agree to abide by this policy.*

Cadmium	mg/L	Below Reporting Limit	4.0E-5	0.0004	EPA 200.8	No evaluation possible
Calcium	mg/L	<b>2.46</b>	0.085	0.2	EPA 200.7	No benchmark available
Carbonate	mg/L	N/A			Carbonate	No benchmark available
Chloride	mg/L	Below Reporting Limit	1.18	2.5	EPA 300.0	No evaluation possible
Chloride-to-Sulfate Mass Ratio		<b>0</b>			CSMR	No benchmark available
Chromium (Total)	mg/L	Below Reporting Limit	0.00011	0.002	EPA 200.8	No evaluation possible
Cobalt	mg/L	Below Reporting Limit	4.0E-5	0.0004	EPA 200.8	No evaluation possible
Copper	mg/L	Below Reporting Limit	0.00054	0.005	EPA 200.8	No evaluation possible
E. coli	P/A	NOT DETECTED			SM 9223B	No evaluation possible
Fluoride	mg/L	Below Reporting Limit	0.0403	0.5	EPA 300.0	No evaluation possible
Iron	mg/L	Below Reporting Limit	0.072	0.2	EPA 200.7	No evaluation possible
Lead	mg/L	Below Reporting Limit	5.0E-5	0.0004	EPA 200.8	No evaluation possible
Lithium	mg/L	<b>0.00241</b>	0.0001	0.001	EPA 200.7	< HGL
Magnesium	mg/L	<b>1.7</b>	0.0408	0.2	EPA 200.7	No benchmark available
Manganese	mg/L	<b>0.000422</b>	0.00017	0.0004	EPA 200.8	< HGL
Mercury	mg/L	Below Reporting Limit	5.0E-5	0.0008	EPA 200.8	No evaluation possible
Molybdenum	mg/L	Below Reporting Limit	4.0E-5	0.0004	EPA 200.8	No evaluation possible
Nickel	mg/L	Below Reporting Limit	0.00024	0.005	EPA 200.8	No evaluation possible
Nitrate (as N)	mg/L	<b>0.16</b>	0.015	0.125	EPA 300.0	< HGL
Nitrite (as N)	mg/L	Below Reporting Limit	0.05	0.125	EPA 300.0	No evaluation possible
Phosphorus	mg/L	Below Reporting Limit	0.0294	0.2	EPA 200.7	No evaluation possible
Potassium	mg/L	<b>0.417</b>	0.192	0.2	EPA 200.7	No benchmark available
Selenium	mg/L	Below Reporting Limit	0.00021	0.001	EPA 200.8	No evaluation possible
Silica	mg/L	<b>9.94</b>	0.0667	0.214	EPA 200.7	No benchmark available
Silver	mg/L	Below Reporting Limit	8.0E-5	0.0004	EPA 200.8	No evaluation possible
Sodium	mg/L	<b>1.46</b>	0.134	0.2	EPA 200.7	No benchmark available
Specific Conductivity	umhos/cm	<b>37.9</b>	1.77	2	SM 2510 B	No benchmark available
Strontium	mg/L	Below Reporting Limit	0.00505	0.05	EPA 200.7	No evaluation possible
Sulfate	mg/L	<b>3.88</b>	1.62	2.5	EPA 300.0	< HGL
Thallium	mg/L	Below Reporting Limit	3.0E-5	0.0004	EPA 200.8	No evaluation possible
Tin	mg/L	Below Reporting Limit	1.0E-5	0.0004	EPA 200.8	No evaluation possible
Titanium	mg/L	Below Reporting Limit	0.00082	0.002	EPA 200.7	No evaluation possible
Total Coliform	P/A	NOT DETECTED			SM 9223B	No evaluation possible
Total HPC	CFU/100mL	Below Reporting Limit	1	200	SM 9215 E	No evaluation possible
Uranium	mg/L	Below Reporting Limit	5.0E-5	0.0004	EPA 200.8	No evaluation possible
Vanadium	mg/L	Below Reporting Limit	0.00036	0.002	EPA 200.8	No evaluation possible
Zinc	mg/L	Below Reporting Limit	0.00428	0.01	EPA 200.7	No evaluation possible

### How To Read Your SimpleLab PDF Report

Your results are being evaluated against the Health Guidance Level.

This is a health protective, non-enforceable drinking water benchmark. HGL is based on the most protective human health benchmark used among public health agencies for a contaminant. Drinking water at or near the HGL over the course of your lifetime is thought to be safe.

**MDL:** The Method Detection Limit is the lowest concentration which the analysis team and instrumentation is configured to measure.

**RL:** The Method Reporting Limit is the lowest level at which the laboratory can confidently and accurately quantify the concentration.

## How To Read Your SimpleLab PDF Report

Your results are being evaluated against the Health Guidance Level.

This is a health protective, non-enforceable drinking water benchmark. HGL is based on the most protective human health benchmark used among public health agencies for a contaminant. Drinking water at or near the HGL over the course of your lifetime is thought to be safe.

MDL: Method Detection Limit. MDL is the lowest concentration of an analyte which testing instrumentation and the analysis team are configured to measure.



## Did you know?

This Tap Score report is easier to understand when viewed online. Access in-depth information about every detection, including health risks and treatment solutions.

[gosimplelab.com/signin](http://gosimplelab.com/signin)